
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001574
12:58 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:06
STERILIZING 3:07
136.5 C 224kPa 3:07
136.4 C 226kPa 3:37
135.9 C 220kPa 4:07
135.9 C 222kPa 4:37
136.4 C 224kPa 5:07
136.2 C 223kPa 5:37
136.1 C 222kPa 6:07
136.1 C 220kPa 6:37
136.1 C 220kPa 6:37
VENTING 6:37
AIR DRYING 7:26
CYCLE COMPLETE 10:55

Digital Signature #
C6F0BD674B41E02B
