
Midmark M11 - v1.0.5
Total Cycles: 2506

Name: RITTER2
Sterilizer ID: V2313228
06/13/2023 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 186.9 F 3.4
4:00 206.1 F 8.9
6:00 222.4 F 8.6
8:00 234.0 F 10.7
10:00 234.7 F 8.5
12:00 239.9 F 10.3
14:00 247.7 F 14.7
16:00 260.4 F 21.8
18:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.3 F 29.4
2:00 272.3 F 29.9
3:00 271.8 F 29.2
4:00 272.6 F 29.9

Min 271.5 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 221.7 F 0.0
10:00 243.6 F 0.0
15:00 233.0 F 0.0
20:00 225.6 F 0.0
25:00 225.9 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

06/13/2023 10:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____