
Midmark M11 - v1.0.5
Total Cycles: 2792

Name: RITTER2
Sterilizer ID: V2313228
08/18/2023 05:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.6 F 0.0
2:00 153.5 F 0.0
4:00 183.1 F 0.0
6:00 209.2 F 1.9
8:00 227.0 F 6.5
10:00 233.8 F 11.3
12:00 233.9 F 8.5
14:00 240.5 F 10.8
16:00 245.6 F 12.8
18:00 245.9 F 13.5
20:00 259.0 F 20.6
22:00 270.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.1 F 28.9
2:00 272.8 F 29.4
3:00 271.6 F 28.9
4:00 272.3 F 29.0

Min 271.5 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 195.4 F 0.0
10:00 199.4 F 0.0
15:00 196.2 F 0.0
20:00 192.6 F 0.0
25:00 188.9 F 0.0
30:00 186.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:46
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:39
CYCLE COMPLETE

08/18/2023 06:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____