
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000012
11:24 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:25
STERILIZING 3:26
136.0 C 220kPa 3:26
136.2 C 223kPa 3:56
136.4 C 224kPa 4:26
136.4 C 224kPa 4:56
136.4 C 224kPa 5:26
136.4 C 224kPa 5:56
136.4 C 224kPa 6:26
136.2 C 223kPa 6:56
136.2 C 223kPa 6:56
VENTING 6:56
AIR DRYING 7:46
CYCLE COMPLETE 7:51

Digital Signature #
EB62307BF47B40A4
