
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002179
10:14 29/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:23
STERILIZING 3:30
136.4 C 225kPa 3:30
136.4 C 225kPa 4:00
136.4 C 225kPa 4:30
136.4 C 222kPa 5:00
136.4 C 225kPa 5:30
136.4 C 226kPa 6:00
136.4 C 223kPa 6:30
136.4 C 226kPa 7:00
136.4 C 226kPa 7:00
VENTING 7:00
AIR DRYING 7:50
CYCLE COMPLETE 36:20

Digital Signature #
0AB1D78179583EA3
