
Midmark M11 - v1.0.0
Total Cycles: 809

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2021 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 179.8 F 3.1
4:00 193.9 F 7.9
6:00 218.9 F 10.0
8:00 228.3 F 8.5
10:00 240.8 F 11.4
12:00 243.9 F 11.9
14:00 241.8 F 10.9
16:00 256.0 F 18.1
18:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 272.7 F 28.6
2:00 271.6 F 28.0
3:00 272.4 F 28.3
4:00 271.9 F 28.4

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 231.6 F 0.0
10:00 225.3 F 0.0
15:00 181.9 F 0.0
20:00 165.4 F 0.0
25:00 157.2 F 0.0
Dry Cycle Aborted

FILLING: 2:07
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:03
DRYING: 27:03
TOTAL CYCLE: 00:55:21

04/26/2021 01:58 PM
ACCEPTED: X
REJECTED:
Operator: U01