
Midmark M11 - v1.0.0
Total Cycles: 1507

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 79.9 F 0.0
4:00 97.0 F 0.0
6:00 130.5 F 1.4
8:00 163.2 F 4.2
10:00 193.9 F 7.4
12:00 212.4 F 10.9
14:00 219.0 F 3.9
16:00 215.2 F 5.8
18:00 226.1 F 9.6
20:00 236.1 F 11.6
22:00 240.3 F 10.0
24:00 238.2 F 8.6
26:00 247.7 F 13.5
28:00 258.6 F 19.4
30:00 268.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.9 F 28.2
2:00 272.0 F 28.4
3:00 271.9 F 28.4
4:00 271.9 F 28.4

Min 271.6 F 28.0
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 233.9 F 0.0
10:00 224.8 F 0.0
15:00 221.7 F 0.0
20:00 222.6 F 0.0
25:00 232.8 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:53
CYCLE COMPLETE

01/11/2022 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____