

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4531

Name: RITTER 1  
Sterilizer ID: V2178198  
05/07/2024 12:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.0 F 0.0  
2:00 104.7 F 0.5  
4:00 123.3 F 2.7  
6:00 151.4 F 5.4  
8:00 177.8 F 8.1  
10:00 194.7 F 11.7  
12:00 209.8 F 4.2  
14:00 217.2 F 8.0  
16:00 228.1 F 12.0  
18:00 229.1 F 7.5  
20:00 240.4 F 12.4  
22:00 238.3 F 10.1  
24:00 250.2 F 15.8  
26:00 260.7 F 21.7  
28:00 270.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.7 F 29.5  
2:00 271.6 F 28.9  
3:00 272.7 F 29.6  
4:00 272.1 F 29.1

Min 271.5 F 28.9  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 205.4 F 0.0  
10:00 207.8 F 0.0  
15:00 224.8 F 0.0  
20:00 222.3 F 0.0  
25:00 218.9 F 0.0  
30:00 231.2 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 29:38  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:08:18  
CYCLE COMPLETE

05/07/2024 01:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_