

-----  
Midmark M11 - v1.0.5  
Total Cycles: 154

Name: RITTER2  
Sterilizer ID: V2313228  
06/29/2021 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.6 F 0.0  
2:00 124.7 F 1.1  
4:00 144.8 F 3.4  
6:00 161.7 F 6.3  
8:00 186.6 F 9.6  
10:00 201.9 F 13.1  
12:00 212.4 F 4.7  
14:00 223.5 F 8.5  
16:00 234.7 F 12.6  
18:00 236.9 F 10.4  
20:00 237.5 F 9.3  
22:00 248.2 F 14.9  
24:00 258.8 F 20.8  
26:00 268.4 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.3 F 29.9  
2:00 271.4 F 29.3  
3:00 272.0 F 30.2  
4:00 272.6 F 30.1

Min 271.0 F 29.1  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.1 F 0.0  
5:00 193.1 F 0.0  
10:00 186.6 F 0.0  
15:00 189.9 F 0.0  
20:00 192.2 F 0.0  
25:00 194.0 F 0.0  
30:00 194.1 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 27:19  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:05:56  
CYCLE COMPLETE

06/29/2021 12:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_