

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1230

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2021 06:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.6 F 0.0  
2:00 74.0 F 0.0  
4:00 77.9 F 0.0  
6:00 99.0 F 0.1  
8:00 139.2 F 0.1  
10:00 168.7 F 0.3  
12:00 188.9 F 2.7  
14:00 217.9 F 5.6  
16:00 223.4 F 8.7  
18:00 225.0 F 12.2  
20:00 227.9 F 6.6  
22:00 236.9 F 11.0  
24:00 234.0 F 7.6  
26:00 246.5 F 13.3  
28:00 243.8 F 11.7  
30:00 255.3 F 17.7  
32:00 265.7 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.6 F 28.6  
2:00 271.5 F 27.8  
3:00 272.6 F 28.7  
4:00 272.1 F 28.1  
  
Min 271.4 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 224.3 F 0.0  
10:00 202.1 F 0.0  
15:00 186.8 F 0.0  
20:00 195.3 F 0.0  
25:00 207.9 F 0.0  
30:00 217.5 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:49  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:11:57  
CYCLE COMPLETE

10/05/2021 07:48 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01