
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 008139
14:42 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:47
136.4 C 223kPa 4:47
136.3 C 224kPa 5:17
136.4 C 221kPa 5:47
136.4 C 222kPa 6:17
136.4 C 223kPa 6:47
136.2 C 221kPa 7:17
136.3 C 222kPa 7:47
136.4 C 223kPa 8:17
136.4 C 223kPa 8:17
VENTING 8:17
AIR DRYING 9:11
CYCLE COMPLETE 9:14

Digital Signature #
EDD6C3FE8B7824D6
