

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4139

Name: RITTER2  
Sterilizer ID: V2313228  
11/20/2024 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 70.2 F 0.0  
4:00 75.2 F 0.0  
6:00 102.0 F 0.1  
8:00 144.0 F 0.2  
10:00 167.3 F 0.5  
12:00 195.7 F 2.1  
14:00 219.1 F 5.5  
16:00 230.8 F 9.1  
18:00 234.9 F 13.1  
20:00 234.1 F 9.1  
22:00 239.8 F 11.2  
24:00 243.7 F 12.5  
26:00 242.3 F 11.6  
28:00 254.9 F 18.0  
30:00 265.9 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.7 F 29.2  
2:00 271.4 F 28.5  
3:00 272.7 F 29.4  
4:00 271.9 F 28.6  
  
Min 271.3 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
5:00 219.4 F 0.0  
10:00 202.3 F 0.0  
15:00 187.1 F 0.0  
20:00 187.9 F 0.0  
25:00 186.2 F 0.0  
30:00 176.8 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:48  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:59  
CYCLE COMPLETE

11/20/2024 11:30 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_