
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001349
9:53 16/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:39
STERILIZING 4:53
136.3 C 223kPa 4:53
136.2 C 224kPa 5:23
136.1 C 222kPa 5:53
136.4 C 225kPa 6:23
136.2 C 223kPa 6:53
136.4 C 223kPa 7:23
136.3 C 221kPa 7:53
136.3 C 224kPa 8:23
136.3 C 224kPa 8:23
VENTING 8:23
AIR DRYING 9:17
CYCLE COMPLETE 24:17

Digital Signature #
8C861CBA800CE141
