
Midmark M11 - v1.0.0
Total Cycles: 3595

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2023 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.1 F 0.0
2:00 95.0 F 0.0
4:00 110.9 F 0.0
6:00 150.2 F 0.0
8:00 175.1 F 0.5
10:00 197.1 F 2.9
12:00 209.1 F 5.6
14:00 216.0 F 8.9
16:00 225.8 F 12.5
18:00 228.9 F 7.3
20:00 239.5 F 11.8
22:00 236.4 F 9.0
24:00 241.7 F 10.9
26:00 246.3 F 13.1
28:00 257.5 F 19.0
30:00 267.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 272.2 F 28.6
2:00 272.0 F 28.2
3:00 271.6 F 28.2
4:00 272.5 F 28.6

Min 271.3 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 206.7 F 0.0
10:00 195.7 F 0.0
15:00 201.2 F 0.0
20:00 208.4 F 0.0
25:00 215.5 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:09:46
CYCLE COMPLETE

07/31/2023 03:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____