
Midmark M11 - v1.0.5
Total Cycles: 2040

Name: RITTER2
Sterilizer ID: V2313228
02/13/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.8 F 0.0
2:00 192.1 F 3.1
4:00 197.7 F 7.9
6:00 217.6 F 13.5
8:00 228.6 F 9.5
10:00 234.4 F 9.1
12:00 243.3 F 12.6
14:00 244.5 F 13.2
16:00 257.6 F 20.2
18:00 269.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.5 F 29.8
2:00 271.5 F 29.1
3:00 272.5 F 30.0
4:00 272.0 F 29.4

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 221.7 F 0.0
10:00 240.4 F 0.0
15:00 237.5 F 0.0
20:00 233.3 F 0.0
25:00 226.7 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

02/13/2023 12:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____