
Midmark M11 - v1.0.0
Total Cycles: 1117

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2021 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 190.6 F 3.3
4:00 218.6 F 8.8
6:00 228.5 F 8.5
8:00 238.4 F 10.9
10:00 234.7 F 8.3
12:00 240.4 F 10.0
14:00 248.2 F 14.1
16:00 260.9 F 21.2
18:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.3 F 28.4
2:00 272.5 F 28.9
3:00 271.7 F 28.1
4:00 272.7 F 28.9

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.3 F 0.0
5:00 201.2 F 0.0
10:00 200.8 F 0.0
15:00 200.5 F 0.0
20:00 195.5 F 0.0
25:00 189.8 F 0.0
30:00 187.0 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

08/25/2021 10:29 AM
ACCEPTED: X
REJECTED:
Operator: U02