

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3210

Name: RITTER 1  
Sterilizer ID: V2178198  
05/10/2023 10:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.3 F 0.0  
2:00 160.7 F 0.4  
4:00 182.2 F 3.7  
6:00 202.6 F 8.3  
8:00 217.7 F 13.5  
10:00 228.0 F 7.9  
12:00 239.4 F 12.8  
14:00 237.5 F 9.8  
16:00 238.4 F 8.9  
18:00 247.8 F 14.0  
20:00 259.6 F 20.5  
22:00 270.1 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.1 F 28.5  
2:00 272.5 F 29.0  
3:00 271.5 F 28.2  
4:00 272.4 F 28.8

Min 271.3 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 235.5 F 0.0  
10:00 237.8 F 0.0  
15:00 225.1 F 0.0  
20:00 221.1 F 0.0  
25:00 215.2 F 0.0  
30:00 210.7 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 22:50  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

05/10/2023 11:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_