
Midmark M11 - v1.0.0
Total Cycles: 2216

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.9 F 0.0
2:00 193.9 F 4.4
4:00 199.4 F 9.2
6:00 211.9 F 8.3
8:00 226.5 F 9.0
10:00 238.4 F 10.0
12:00 239.2 F 11.4
14:00 237.8 F 9.5
16:00 250.7 F 16.2
18:00 262.6 F 23.0
20:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.9 F 29.5
2:00 271.8 F 28.8
3:00 272.9 F 29.7
4:00 272.5 F 29.2

Min 271.6 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 232.7 F 0.0
10:00 239.7 F 0.0
15:00 231.0 F 0.0
20:00 229.6 F 0.0
25:00 228.3 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

08/09/2022 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____