
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007757
14:51 15/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:03
STERILIZING 2:57
136.0 C 220kPa 2:57
136.4 C 225kPa 3:27
136.4 C 225kPa 3:57
136.2 C 223kPa 4:27
136.4 C 224kPa 4:57
136.3 C 223kPa 5:27
136.4 C 222kPa 5:57
136.2 C 217kPa 6:27
136.2 C 217kPa 6:27
VENTING 6:27
AIR DRYING 7:18
CYCLE COMPLETE 13:56

Digital Signature #
2A39040D765B8ED1
