
Midmark M11 - v1.0.0
Total Cycles: 2694

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2022 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.3 F 0.0
2:00 196.8 F 3.9
4:00 221.4 F 9.8
6:00 227.6 F 6.4
8:00 238.4 F 11.9
10:00 237.9 F 9.9
12:00 238.0 F 8.3
14:00 251.7 F 16.0
16:00 264.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.4 F 28.2
3:00 272.6 F 28.9
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.6 F 0.0
5:00 229.8 F 0.0
10:00 235.3 F 0.0
15:00 229.5 F 0.0
20:00 228.4 F 0.0
25:00 229.5 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 2:54
HEATING: 17:52
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

12/19/2022 05:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____