

-----  
Midmark M11 - v1.0.0  
Total Cycles: 887

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2021 06:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.0 F 0.0  
2:00 78.2 F 0.0  
4:00 91.3 F 0.0  
6:00 130.4 F 0.1  
8:00 159.5 F 0.2  
10:00 178.6 F 0.3  
12:00 219.3 F 3.7  
14:00 226.4 F 6.9  
16:00 235.2 F 10.2  
18:00 236.7 F 9.9  
20:00 239.9 F 10.4  
22:00 234.8 F 7.9  
24:00 246.7 F 13.5  
26:00 243.8 F 12.0  
28:00 256.8 F 18.7  
30:00 268.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 271.9 F 28.1  
2:00 272.4 F 28.5  
3:00 272.4 F 28.8  
4:00 271.5 F 28.0

Min 271.5 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
5:00 232.7 F 0.0  
10:00 217.8 F 0.0  
15:00 216.6 F 0.0  
20:00 216.8 F 0.0  
25:00 223.7 F 0.0  
Dry Cycle Aborted

FILLING: 0:50  
HEATING: 31:13  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 28:17  
TOTAL CYCLE: 01:07:17

06/14/2021 07:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01