

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3226

Name: RITTER2  
Sterilizer ID: V2313228  
12/18/2023 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.5 F 0.0  
2:00 74.7 F 0.0  
4:00 78.2 F 0.1  
6:00 107.7 F 1.5  
8:00 146.7 F 4.7  
10:00 170.1 F 8.0  
12:00 218.6 F 11.9  
14:00 214.8 F 4.8  
16:00 220.8 F 8.3  
18:00 231.2 F 12.2  
20:00 231.1 F 7.7  
22:00 242.1 F 12.6  
24:00 239.1 F 10.5  
26:00 251.2 F 16.5  
28:00 261.8 F 22.6  
30:00 271.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.6  
2:00 271.5 F 29.4  
3:00 272.2 F 29.3  
4:00 271.9 F 29.7

Min 271.3 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
5:00 221.7 F 0.0  
10:00 225.7 F 0.0  
15:00 213.8 F 0.0  
20:00 207.3 F 0.0  
25:00 207.3 F 0.0  
30:00 214.9 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:08:37  
CYCLE COMPLETE

12/18/2023 10:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_