

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1032

Name: RITTER 1  
Sterilizer ID: V2178198  
07/20/2021 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
2:00 190.2 F 3.1  
4:00 187.4 F 6.6  
6:00 211.5 F 11.5  
8:00 222.5 F 5.5  
10:00 231.2 F 10.3  
12:00 236.1 F 7.9  
14:00 243.3 F 12.9  
16:00 242.0 F 11.4  
18:00 254.9 F 18.0  
20:00 266.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.1 F 29.1  
2:00 272.1 F 28.7  
3:00 271.8 F 28.9  
4:00 272.4 F 29.0  
  
Min 271.3 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 224.1 F 0.0  
10:00 226.6 F 0.0  
15:00 230.7 F 0.0  
20:00 224.3 F 0.0  
25:00 221.4 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 21:30  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:53  
CYCLE COMPLETE

07/20/2021 10:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01