
Midmark M11 - v1.0.0
Total Cycles: 3744

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2023 02:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.0 F 0.0
2:00 176.8 F 2.9
4:00 200.5 F 7.6
6:00 216.6 F 12.9
8:00 226.5 F 7.9
10:00 238.8 F 13.1
12:00 238.9 F 10.7
14:00 236.5 F 8.9
16:00 250.6 F 15.7
18:00 262.0 F 22.2
20:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.8
2:00 271.8 F 28.8
3:00 272.2 F 28.6
4:00 271.8 F 28.8

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 227.9 F 0.0
10:00 232.8 F 0.0
15:00 229.1 F 0.0
20:00 226.9 F 0.0
25:00 229.8 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 3:07
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

09/08/2023 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____