
Midmark M11 - v1.0.0
Total Cycles: 1444

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 182.2 F 3.5
4:00 196.0 F 7.8
6:00 218.8 F 11.8
8:00 229.6 F 8.1
10:00 241.3 F 12.1
12:00 242.0 F 11.5
14:00 239.6 F 10.0
16:00 253.8 F 17.0
18:00 265.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.1 F 28.4
2:00 271.8 F 28.5
3:00 272.1 F 28.3
4:00 271.8 F 28.5

Min 271.2 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 218.7 F 0.0
10:00 222.4 F 0.0
15:00 224.7 F 0.0
20:00 227.4 F 0.0
25:00 210.6 F 0.0
30:00 204.8 F 0.0

Drying Complete

FILLING: 2:13
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

12/09/2021 03:52 PM
ACCEPTED: X
REJECTED:
Operator: U02