
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000743
13:23 30/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:31
STERILIZING 3:30
136.5 C 224kPa 3:30
136.2 C 223kPa 4:01
136.1 C 220kPa 4:31
136.1 C 219kPa 5:01
136.4 C 224kPa 5:31
136.4 C 225kPa 6:01
136.4 C 225kPa 6:31
136.4 C 224kPa 7:01
136.4 C 224kPa 7:01
VENTING 7:01
AIR DRYING 7:50
CYCLE COMPLETE 7:53

Digital Signature #
490EFCEBCCA0C24B
