

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3898

Name: RITTER2  
Sterilizer ID: V2313228  
08/28/2024 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.7 F 0.0  
2:00 71.3 F 0.0  
4:00 84.7 F 0.0  
6:00 113.7 F 0.5  
8:00 154.5 F 2.9  
10:00 176.6 F 6.4  
12:00 194.1 F 10.4  
14:00 209.7 F 6.7  
16:00 219.1 F 6.7  
18:00 229.5 F 10.6  
20:00 232.4 F 8.1  
22:00 241.5 F 12.2  
24:00 240.3 F 10.8  
26:00 253.0 F 17.2  
28:00 264.0 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.1 F 28.8  
2:00 272.6 F 29.2  
3:00 272.6 F 29.6  
4:00 271.7 F 28.8

Min 271.6 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.2 F 0.0  
5:00 208.6 F 0.0  
10:00 222.5 F 0.0  
15:00 217.8 F 0.0  
20:00 214.4 F 0.0  
25:00 213.7 F 0.0  
30:00 213.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 29:57  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:00  
CYCLE COMPLETE

08/28/2024 09:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_