
Midmark M11 - v1.0.5
Total Cycles: 1188

Name: RITTER2
Sterilizer ID: V2313228
06/15/2022 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 173.7 F 2.3
4:00 178.8 F 5.7
6:00 198.5 F 10.4
8:00 214.2 F 5.7
10:00 228.0 F 10.3
12:00 230.1 F 7.8
14:00 242.1 F 13.0
16:00 241.6 F 11.7
18:00 254.2 F 18.1
20:00 265.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.1 F 29.8
2:00 272.7 F 29.6
3:00 273.1 F 29.9
4:00 272.0 F 29.3

Min 270.9 F 28.8
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 219.3 F 0.0
10:00 226.2 F 0.0
15:00 224.7 F 0.0
20:00 220.2 F 0.0
25:00 219.4 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

06/15/2022 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____