
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008062
8:42 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:31
STERILIZING 4:26
136.4 C 220kPa 4:26
136.1 C 222kPa 4:56
136.5 C 221kPa 5:26
136.5 C 224kPa 5:56
136.3 C 221kPa 6:26
136.4 C 223kPa 6:56
136.3 C 220kPa 7:26
136.3 C 220kPa 7:56
136.3 C 220kPa 7:56
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 30:17

Digital Signature #
C335F4AA12120A55
