
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001743
12:32 17/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 3:01
136.4 C 226kPa 3:01
136.3 C 225kPa 3:31
136.4 C 224kPa 4:01
136.2 C 225kPa 4:31
136.0 C 223kPa 5:01
136.4 C 227kPa 5:31
136.3 C 225kPa 6:01
136.3 C 225kPa 6:31
136.3 C 225kPa 6:31
VENTING 6:31
AIR DRYING 7:23
CYCLE COMPLETE 12:10

Digital Signature #
C291BF5F206B7105
