

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2989

Name: RITTER 1  
Sterilizer ID: V2178198  
03/09/2023 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
2:00 192.5 F 4.4  
4:00 216.0 F 9.4  
6:00 221.8 F 6.9  
8:00 235.0 F 9.7  
10:00 234.0 F 8.3  
12:00 245.2 F 13.3  
14:00 244.0 F 12.2  
16:00 257.3 F 19.2  
18:00 269.0 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 28.9  
2:00 271.5 F 28.1  
3:00 272.5 F 29.1  
4:00 272.3 F 28.6  
  
Min 271.2 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 232.2 F 0.0  
10:00 232.1 F 0.0  
15:00 229.3 F 0.0  
20:00 228.0 F 0.0  
25:00 229.2 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 19:05  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:05  
CYCLE COMPLETE

03/09/2023 02:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_