
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000399
14:53 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:23
STERILIZING 4:23
136.0 C 230kPa 4:23
136.4 C 233kPa 4:53
136.3 C 233kPa 5:23
136.3 C 232kPa 5:53
136.3 C 233kPa 6:23
136.4 C 233kPa 6:53
136.0 C 230kPa 7:23
136.4 C 233kPa 7:53
136.4 C 233kPa 7:53
VENTING 7:54
AIR DRYING 8:48
CYCLE COMPLETE 8:50

Digital Signature #
79713E1A2E1B93D7
