
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004560
13:14 18/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:56
STERILIZING 2:55
136.5 C 221kPa 2:55
136.4 C 223kPa 3:25
136.4 C 224kPa 3:55
136.4 C 225kPa 4:25
136.4 C 224kPa 4:55
136.4 C 223kPa 5:25
136.3 C 223kPa 5:55
136.4 C 224kPa 6:25
136.4 C 224kPa 6:25
VENTING 6:25
AIR DRYING 7:15
CYCLE COMPLETE 7:24

Digital Signature #
151D767B85977AA3
