
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000861
11:17 13/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:56
STERILIZING 4:10
136.0 C 220kPa 4:10
136.4 C 232kPa 4:40
136.4 C 233kPa 5:10
136.1 C 231kPa 5:40
136.1 C 230kPa 6:10
136.4 C 233kPa 6:40
136.4 C 232kPa 7:10
136.4 C 233kPa 7:40
136.4 C 233kPa 7:40
VENTING 7:40
AIR DRYING 8:31
CYCLE COMPLETE 19:05

Digital Signature #
98914FAD9370C3D1
