
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002085
7:17 21/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:21
PRESSURIZING 1:27
STERILIZING 2:33
136.4 C 221kPa 2:33
136.4 C 223kPa 3:03
136.4 C 222kPa 3:33
136.4 C 221kPa 4:03
136.0 C 219kPa 4:33
136.3 C 222kPa 5:03
136.2 C 221kPa 5:33
136.4 C 223kPa 6:03
136.4 C 223kPa 6:03
VENTING 6:03
AIR DRYING 6:55
CYCLE COMPLETE 6:56

Digital Signature #
7EA92ACC2EA9DBC7
