

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4700

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2024 11:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.4 F 0.0  
2:00 133.6 F 1.8  
4:00 143.6 F 4.5  
6:00 167.7 F 7.7  
8:00 195.8 F 11.0  
10:00 211.3 F 4.8  
12:00 216.2 F 7.5  
14:00 229.2 F 11.5  
16:00 227.8 F 6.8  
18:00 239.9 F 11.9  
20:00 237.0 F 9.6  
22:00 249.2 F 15.2  
24:00 259.8 F 21.0  
26:00 269.3 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.1 F 29.4  
2:00 272.9 F 29.4  
3:00 272.1 F 29.1  
4:00 271.4 F 28.7

Min 271.2 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 210.4 F 0.0  
10:00 228.2 F 0.0  
15:00 227.9 F 0.0  
20:00 225.0 F 0.0  
25:00 224.5 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 27:14  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:06:01  
CYCLE COMPLETE

07/05/2024 12:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_