

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1484

Name: RITTER2  
Sterilizer ID: V2313228  
09/13/2022 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
2:00 186.6 F 3.3  
4:00 192.4 F 8.4  
6:00 214.5 F 9.7  
8:00 229.6 F 10.0  
10:00 235.8 F 8.6  
12:00 241.7 F 12.0  
14:00 244.6 F 13.5  
16:00 257.7 F 20.6  
18:00 269.4 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.5  
1:00 272.7 F 30.2  
2:00 271.8 F 29.4  
3:00 273.1 F 30.4  
4:00 272.4 F 29.8  
  
Min 271.8 F 29.4  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 211.9 F 0.0  
10:00 220.8 F 0.0  
15:00 224.5 F 0.0  
20:00 225.4 F 0.0  
25:00 225.7 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 3:03  
HEATING: 19:01  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

09/13/2022 12:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_