

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2074

Name: RITTER2  
Sterilizer ID: V2313228  
02/22/2023 09:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.1 F 0.0  
2:00 154.4 F 0.0  
4:00 163.4 F 0.3  
6:00 187.9 F 4.2  
8:00 214.0 F 8.8  
10:00 227.0 F 9.5  
12:00 235.0 F 10.3  
14:00 235.0 F 8.1  
16:00 241.8 F 11.4  
18:00 246.3 F 13.7  
20:00 258.9 F 20.6  
22:00 270.3 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.4  
2:00 271.6 F 29.0  
3:00 272.5 F 29.5  
4:00 271.5 F 28.8

Min 271.4 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 207.5 F 0.0  
10:00 199.2 F 0.0  
15:00 201.1 F 0.0  
20:00 202.5 F 0.0  
25:00 228.9 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 22:51  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:52  
CYCLE COMPLETE

02/22/2023 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_