
Midmark M11 - v1.0.0
Total Cycles: 3582

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2023 04:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 176.5 F 0.3
4:00 197.2 F 3.9
6:00 213.0 F 7.4
8:00 224.3 F 11.7
10:00 228.5 F 7.1
12:00 239.4 F 12.1
14:00 237.9 F 9.8
16:00 239.1 F 9.2
18:00 249.4 F 14.8
20:00 260.8 F 21.1
22:00 271.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.3 F 28.5
2:00 272.1 F 28.7
3:00 271.9 F 28.3
4:00 272.2 F 28.8

Min 271.5 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 204.9 F 0.0
10:00 206.7 F 0.0
15:00 207.1 F 0.0
20:00 207.6 F 0.0
25:00 210.5 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

07/25/2023 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____