

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1225

Name: RITTER2  
Sterilizer ID: V2313228  
06/27/2022 08:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.5 F 0.0  
2:00 188.2 F 2.4  
4:00 188.2 F 6.7  
6:00 208.5 F 12.3  
8:00 223.0 F 7.0  
10:00 236.0 F 12.3  
12:00 236.4 F 9.8  
14:00 237.6 F 9.2  
16:00 248.8 F 14.9  
18:00 260.6 F 21.5  
20:00 271.3 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.5 F 29.2  
2:00 271.4 F 28.6  
3:00 272.7 F 29.4  
4:00 271.8 F 28.7

Min 271.4 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 206.0 F 0.0  
10:00 220.2 F 0.0  
15:00 224.0 F 0.0  
20:00 222.3 F 0.0  
25:00 221.0 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:49  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:59:31  
CYCLE COMPLETE

06/27/2022 09:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_