

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002064  
9:19 19/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:51  
PRESSURIZING 3:51  
STERILIZING 4:55  
136.0 C 220kPa 4:55  
136.4 C 224kPa 5:25  
136.4 C 223kPa 5:55  
136.1 C 222kPa 6:25  
136.4 C 223kPa 6:55  
136.2 C 223kPa 7:25  
136.4 C 223kPa 7:55  
136.3 C 223kPa 8:25  
136.3 C 223kPa 8:25  
VENTING 8:26  
AIR DRYING 9:20  
CYCLE COMPLETE 9:23

-----  
Digital Signature #  
4BBC5DE57881A423  
-----