
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002747
15:54 29/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:34
STERILIZING 4:31
136.1 C 221kPa 4:31
136.3 C 222kPa 5:01
136.4 C 223kPa 5:31
136.4 C 225kPa 6:01
136.2 C 220kPa 6:31
136.4 C 225kPa 7:01
136.4 C 225kPa 7:31
136.2 C 224kPa 8:01
136.2 C 224kPa 8:01
VENTING 8:01
AIR DRYING 8:54
CYCLE COMPLETE 9:05

Digital Signature #
510A9CEFED4ADDF6
