
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006609
10:47 14/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:15
STERILIZING 3:13
136.4 C 222kPa 3:13
136.1 C 223kPa 3:43
136.4 C 224kPa 4:13
136.4 C 224kPa 4:43
136.2 C 224kPa 5:13
136.4 C 223kPa 5:43
136.3 C 223kPa 6:13
136.4 C 224kPa 6:43
136.4 C 224kPa 6:43
VENTING 6:43
AIR DRYING 7:34
CYCLE COMPLETE 7:41

Digital Signature #
FB3CFB88CAB3628F
