

-----  
Midmark M11 - v1.0.5  
Total Cycles: 454

Name: RITTER2  
Sterilizer ID: V2313228  
10/14/2021 02:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.6 F 0.0  
2:00 164.3 F 3.3  
4:00 187.0 F 8.0  
6:00 211.0 F 10.7  
8:00 225.6 F 9.1  
10:00 234.8 F 9.6  
12:00 243.9 F 13.4  
14:00 244.5 F 13.0  
16:00 258.0 F 20.3  
18:00 270.1 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.6 F 29.7  
2:00 271.5 F 28.9  
3:00 272.4 F 30.0  
4:00 272.5 F 29.5  
  
Min 271.2 F 28.9  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 222.8 F 0.0  
10:00 229.8 F 0.0  
15:00 223.7 F 0.0  
20:00 222.5 F 0.0  
25:00 230.8 F 0.0  
30:00 216.6 F 0.0  
Drying Complete

FILLING: 2:29  
HEATING: 18:54  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:29  
CYCLE COMPLETE

10/14/2021 03:03 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02