
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007888
8:10 26/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:24
STERILIZING 3:24
136.0 C 220kPa 3:24
136.4 C 223kPa 3:54
136.4 C 223kPa 4:24
136.4 C 223kPa 4:54
135.9 C 218kPa 5:24
136.3 C 223kPa 5:54
136.4 C 221kPa 6:24
136.4 C 223kPa 6:54
136.4 C 223kPa 6:54
VENTING 6:54
AIR DRYING 7:43
CYCLE COMPLETE 22:52

Digital Signature #
BB837674598E3258
