
Midmark M11 - v1.0.0
Total Cycles: 3913

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2023 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.3 F 0.0
2:00 153.9 F 3.0
4:00 175.2 F 7.2
6:00 198.5 F 12.1
8:00 222.7 F 6.2
10:00 234.9 F 11.4
12:00 233.8 F 8.2
14:00 242.3 F 11.3
16:00 245.1 F 12.8
18:00 257.7 F 19.6
20:00 269.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.7 F 29.0
2:00 271.1 F 28.4
3:00 272.6 F 29.1
4:00 271.6 F 28.5

Min 271.0 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 199.0 F 0.0
10:00 202.7 F 0.0
15:00 201.6 F 0.0
20:00 198.9 F 0.0
25:00 196.2 F 0.0
30:00 193.9 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

10/17/2023 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____