

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3841

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2024 12:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.6 F 0.0  
2:00 161.7 F 2.3  
4:00 172.4 F 5.9  
6:00 195.9 F 10.4  
8:00 213.4 F 5.4  
10:00 224.7 F 9.4  
12:00 233.2 F 9.7  
14:00 240.2 F 12.3  
16:00 240.1 F 11.2  
18:00 253.6 F 17.7  
20:00 264.7 F 24.4  
22:00 271.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 273.0 F 30.2  
2:00 271.8 F 29.7  
3:00 272.4 F 29.6  
4:00 272.6 F 30.1

Min 271.0 F 28.9  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.2 F 0.0  
5:00 232.3 F 0.0  
10:00 227.3 F 0.0  
15:00 224.4 F 0.0  
20:00 224.8 F 0.0  
25:00 225.0 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:00  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:00:52  
CYCLE COMPLETE

08/08/2024 01:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_