

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4055

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2024 10:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.3 F 0.0  
2:00 119.7 F 0.0  
4:00 135.0 F 0.0  
6:00 169.4 F 1.7  
8:00 187.9 F 5.0  
10:00 202.2 F 8.4  
12:00 214.0 F 12.1  
14:00 221.5 F 6.2  
16:00 230.0 F 10.7  
18:00 233.1 F 8.1  
20:00 243.7 F 13.1  
22:00 242.9 F 12.0  
24:00 254.4 F 18.0  
26:00 264.9 F 24.3  
28:00 271.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.4 F 29.6  
2:00 271.9 F 28.9  
3:00 271.1 F 28.8  
4:00 272.1 F 29.6  
  
Min 271.1 F 28.6  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
5:00 209.6 F 0.0  
10:00 212.2 F 0.0  
15:00 215.6 F 0.0  
20:00 214.0 F 0.0  
25:00 214.9 F 0.0  
30:00 219.0 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 28:03  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:06:46  
CYCLE COMPLETE

10/25/2024 11:25 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_