

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001125  
11:39 20/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:24  
PRESSURIZING 3:28  
STERILIZING 4:33  
136.0 C 220kPa 4:33  
136.4 C 226kPa 5:03  
136.1 C 222kPa 5:33  
136.4 C 225kPa 6:03  
136.2 C 224kPa 6:33  
136.4 C 225kPa 7:03  
136.1 C 223kPa 7:33  
136.4 C 222kPa 8:03  
136.4 C 222kPa 8:03  
VENTING 8:03  
AIR DRYING 8:59  
CYCLE COMPLETE 14:24  
-----

Digital Signature #  
E6286E9E6D15A032  
-----