
Midmark M11 - v1.0.0
Total Cycles: 3083

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2023 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.2 F 0.0
2:00 148.5 F 2.0
4:00 166.8 F 5.0
6:00 194.7 F 9.2
8:00 211.8 F 8.6
10:00 223.4 F 7.8
12:00 235.6 F 12.7
14:00 235.5 F 9.4
16:00 237.8 F 9.3
18:00 247.2 F 13.9
20:00 259.4 F 20.5
22:00 270.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.7
2:00 272.0 F 28.9
3:00 272.0 F 28.5
4:00 272.4 F 29.1

Min 271.4 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 215.8 F 0.0
10:00 222.8 F 0.0
15:00 217.3 F 0.0
20:00 218.1 F 0.0
25:00 217.8 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

04/10/2023 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____